

# Characterization of an $^{241}\text{AmBe}$ neutron irradiation facility by different spectrometric techniques

***E. Gallego<sup>1</sup>, K. Amgarou<sup>2</sup>, R. Bedogni<sup>3</sup>,  
C. Domingo<sup>2</sup>, A. Esposito<sup>3</sup>, A. Lorente<sup>1</sup>, R. Méndez<sup>4</sup>,  
H.R. Vega-Carrillo<sup>5</sup>***

*<sup>1</sup> Universidad Politécnica de Madrid (Spain)*

*<sup>2</sup> Universitat Autònoma de Barcelona (UAB)*

*<sup>3</sup> INFN – LNF (Italy)*

*<sup>4</sup> LMRI CIEMAT (Spain)*

*<sup>5</sup> Universidad Autónoma de Zacatecas (Mexico)*



POLITÉCNICA

UAB

Universitat Autònoma  
de BarcelonaINDUSTRIALES  
ETSII | UPM

Ciemat

Centro de Investigaciones  
Energéticas, Medioambientales  
y Tecnológicas


IRPA13 : Glasgow

13th International Congress  
of the International Radiation  
Protection Association

13 - 18 May 2012

GOBIERNO  
DE ESPAÑAMINISTERIO  
DE ECONOMÍA  
Y COMPETITIVIDAD

- An automated panoramic irradiator with a 111 GBq (3 Ci)  $^{241}\text{Am-Be}$  neutron source is installed in a bunker-type large room (16.25 m long, 8.90 m width, 8 m high) at UPM.
- The irradiation bench is placed at 3 m from the floor and at about 4.5 m from any lateral wall.
- A neutron spectrometry campaign was organized with four groups participating with different Bonner Sphere Spectrometers (BSS) and using different spectral unfolding codes.
- The objective is to better characterize the facility, but also the intercomparison itself.



General view of the facility. The red circle indicates the position of the source. The source operation is fully automated and remote controlled.

**Ludlum-BSS:** Six spheres of high-density polyethylene\* with diameters: 2", 3", 5", 8", 10" and 12".

\*  $d = (0.96 \text{ g/cm}^3)$  Determined from weight and volume measurements

**Central Detector:** Scintillator crystal of  $\text{Li}^6\text{I}(\text{Eu})$  of small size  $0,4 \text{ } \varnothing \times 0,4 \text{ cm}$ .

**Electronics:** Data acquisition: ASA-100, HT= 800 V.  
Pre-amp: ORTEC, mod 109PC Preamplifier, X1  
Software: Genie 2000



**BUNKIUT code with the response matrix UTA 4, 25 energy bins (collapsed from 171).**

**Uncertainties not explicitly addressed**






Fig. 1. The calculated 171-neutron group responses for the 4 mm LiI detector and the detector inside 5.08, 7.62, 12.7, 20.32, 30.48, 38.1 and 45.72 cm diameter polyethylene spheres.






**Spheres: diameters:** bare, 2", 3", 5", 7", 8", 10", 12",  
12"+Pb (8 cm internal diameter, 1 cm lead)  
7"+Pb (4" internal diameter, 1/2" lead)


**Density:** 0.95 g.cm<sup>-3</sup>

**Central detector:** 4x4 <sup>6</sup>LiI(Eu)

**Unfolding method:** FRUIT 3.0 (NIM A 580 1301-1309)



Reproducibility check device





**MC code used: MCNPX 2.4.0**

## **Validation experiments**

BSS response matrix overall uncertainty  $u_{mat}$  ( $\pm 3\%$ )

- 1) Determined with irradiations in continuous reference fields (Am-Be, Cf, Cf(D<sub>2</sub>O), thermal) at ENEA Bologna in 2005 - 2006
- 2) Confirmed with monochromatic beams at JRC-Geel (2 MeV, 5 MeV, 16 MeV) in Jan 2006
- 3) Results of monochromatic beams at PTB, March 2009 (24 keV, 144 keV, 1.2 MeV, 8 MeV, 19 MeV) under elaboration

Installation: 1 Ci Am-Be source (INFN-LNF) calibrated at NPL in 1986 and certified with less than 1% uncertainty.

Method: Shadow-cones with cones specifically designed for this BSS





## FRUIT 3.0 (Nucl. Instr. and Meth. A, 580, 1301-1309. 2007)

**Uncertainty treatment**      **YES**

**Pre-information**              **Based in physical environment related to  
neutron production physics**

**Validation**                      **YES**  
- reference sources  
- GSI comparison (2006)

- |  |                |   |
|--|----------------|---|
| ✓ counting   | $u_c$          | (1%-2%)   |
| ✓ anisotropy about cylindrical axis                | $u_{anis}$     | (3%) determined from UPM report                                   |
| ✓ 0° / 90° anisotropy                              | $u_{90^\circ}$ | (1%) determined from NPL report<br>CIRM 24 for same encapsulation |
| ✓ BSS response matrix overall uncertainty          | $u_{mat}$      | (3%)  |
| ✓ BSS calibration & time stability                 | $u_f$          | (2%)  |
| ✓ Unfolding procedures                             |                |   |
|  | Fluence        | 1% - 2%   |
|  | $H^*(10)$      | 4% - 7%   |
| ✓ fluence-to- $H^*$ average conversion coefficient |                | 4% ISO 8529-2   |

Disregarded: height of the source, measurement distance (~ 0.1%)

## Spheres:

Diameters: 2.5", 3", 4.2", 5", 6", 8", 10", 12", 2.5"+Cd, 3"+Cd, 4.2"+Cd

Polyethylene density:  $0.95 \text{ g}\cdot\text{cm}^{-3}$


## Central detector characteristics:

05NH1 from Eurisys.  $^3\text{He}$  filled proportional counter at 8 kPa pressure.  
cylindrical 9 mm x 10 mm.

## Unfolding method and references

FRUIT 3.0 unfolding code

Nucl. Instr. and Meth. A 580, 1301-1309. 2007



- counting < 1.5%
  - geometry (negligible)
  - anisotropies ~ 2%
  - BSS response matrix (simulation) < 1%
  - BSS calibration ~ 3%
  - BSS time stability 0.2% in 12 h
  - unfolding < 2%
  - Fitting method 2%
- Resulting fluence: < 5%



MC code used: MCNPX 2.4.0 and 2.5.0

Validation experiments:

PTB (monenergetic 250 keV, 565 keV, 1.2 MeV, 2.5 MeV, 5 MeV, 14.8 MeV)

IRSN Cadarache (AmBe, Cf, Cf+D<sub>2</sub>O/Cd, SIGMA)

Uncertainty < 3%


**BSS calibration**


April-May 2006, Am-Be and Cf sources a IRSN Cadarache.

**Routine check of the BSS working point**

Not routinely. Sporadic checks.

AmBe Frascati March 2008





**CIEMAT-BSS:** 12 spheres of high density polyethylene with dimensions: 3", 3.5", 4", 4.5", 5", 6", 7", 8", 9", 9.5", 10" and 12".

**Central detector** is a SP9  $^3\text{He}$  spherical proportional counter  
Pressure:  $228.5 \pm 2.0$  kPa  
Voltage: 800V

**Unfolding method:** UMG 3.3 package (MAXED+GRAVEL)



- Counting uncertainty less than 1%
- Geometry uncertainty is not considered.
- Anisotropies uncertainties
- BSS response matrix uncertainty not directly considered.
- BSS calibration uncertainty not considered too.
- Unfolding uncertainty considered in determination of neutron fluence and  $H^*(10)$  less than 0.5%



## Response function (RF) determined by PTB.


- MC code used: MCNP with corrections for PE density and geometry dimensions of spheres.
- Calibrated at PTB (June, 2007) with a reference  $^{252}\text{Cf}$  source calibrated at NPL.
- Validated using monoenergetic neutrons with energies: 144keV, 565keV, 2.5MeV and 15MeV.




- **UMG3.3** unfolding pack has been employed: **GRAVEL** and **MAXED** consecutively and **IQU** for statistical analysis.
- Input data:
  - CIEMAT-BSS RF
  - Cf spectrum as initial spectrum
  - Measurements
- Output data:
  - Output spectrum expressed in 20 energies by decade
  - Fluence rate
  - $H^*(10)$



- Thermal component practically constant and independent of the source-detector distance.
- For epithermal and fast components the shape of the spectra and their values are quite similar.






Normalized spectra (per unit fluence) obtained by the four groups for 115 cm distance point

$$\dot{\phi} = \int_E \dot{\Phi}_E (E) dE$$

Distance	100 cm	115 cm	150 cm
	<b>Total neutron fluence rate, <math>\Phi</math> (cm<sup>-2</sup>·s<sup>-1</sup>)</b>		
<b>UPM-UAZ</b>	62 ± 2	49 ± 2	33 ± 1
<b>INFN</b>	61 ± 3	49 ± 2	32.8 ± 1.2
<b>UAB</b>	64.1 ± 2.6	49.9 ± 2.0	34.1 ± 1.4
<b>CIEMAT</b>	64.3 ± 0.3	50.1 ± 0.2	31.8 ± 0.1

$$\dot{H}^*(10) = \int_E \dot{\Phi}_E(E) h^*(10) dE$$

Distance	100 cm		115 cm		150 cm	
Ambient dose equivalent rate ( $\mu\text{Sv}\cdot\text{h}^{-1}$ )						
<b>UPM-UAZ</b>	77.5	$\pm 2.3$	59.8	$\pm 1.8$	37.3	$\pm 1.1$
<b>INFN</b>	77	$\pm 6$	61	$\pm 5$	37	$\pm 3$
<b>UAB</b>	80.5	$\pm 5.6$	61.8	$\pm 4.3$	40.0	$\pm 2.8$
<b>CIEMAT</b>	75.9	$\pm 0.3$	57.6	$\pm 0.2$	35.1	$\pm 0.1$
<b>LB-6411 (UPM)</b>	79.5	$\pm 0.6$	61.3	$\pm 0.5$	38.4	$\pm 0.8$



- The study has offered a good opportunity to compare results from a set of different BSS, unfolding tools and experimental teams.
- The results were encouraging, showing a reasonable agreement with regard to the main quantities studied.
- However, the differences encountered should be explained, and the results consolidated.
- Relevant features to be determined are the source strength and its anisotropy.
  - Source strength determination is still work on progress.
  - Source anisotropy has been measured after this study using a device designed for this purpose.
- Monte Carlo calculations are being utilized to get a better understanding of the experimental results.

This work was developed within the framework of the agreement between CIEMAT and UPM, with the sponsoring of the Nuclear Safety Council of Spain (CSN) with the objective of development of a neutron metrology system for Spain.